


NITRILECARE EXAMINATION GLOVES

Specifications & Certifications



Multi-Purpose



Product Technical Data Sheet	
Type	Powder-Free, Examination Glove
Specification	Non-Sterile/Disposable
Cuff	Beaded
Weight	Average 3.5g (0.12 fl oz) +/-0.3g
Colour	Indigo Blue
Primary Material	Nitrile
Surface	External: Finger textured
Powder Control	<=2mg/glove
Packaging	100 pcs/box
Size & Product Code	Small: NG-NTCR-001
	Medium: NG-NTCR-002
	Large: NG-NTCR-003
	Extra-large: NG-NTCR-004
Product Name	Nitrile care Examination Gloves
Brand	Careplus
Application Setting	Medium risk medical, dental, procedures, pathology labs, patient care and food handling.

Certifications



Product Properties				
Test Method	Characteristics		Requirement	Median
ASTM D6319	Dimensions (mm)	Width (M)	95 ± 10	95
		Length (M)	Min 230	240 (9.5")
	Thickness single (mm)	Finger	Min 0.05	0.10 (4.0mils)
		Palm	Min 0.05	0.08 (3.1mils)
ASTM D5151	Freedom from holes		No leakage	NA
ASTM D412	Before accelerated aging.			
	Tensile Strength (MPa)		Min 14	30
	Ultimate elongation (%)		Min 500	650
ASTM D573	After accelerated aging at (70 ± 2° C 166±2 hr).			
	Tensile Strength (MPa)		Min 14	14
	Ultimate elongation (%)		Min 400	400
Palm Width	Small: 80mm +/- 10	Weight	3.5g +/- 0.3g (0.12 fl oz)	
	Medium: 95mm +/- 10	Length	9.5 inch glove	
	Large: 110mm +/- 10	Thickness	3.5 mil glove	
	Extra-large: 120mm +/- 10			
US Standards	ASTM D6319, ASTM F1671 (Viral)			
FDA Information	FDA 510(k) - K172015 Click here for FDA Website			
Quality Standards	ISO 9001:2015, EN ISO 13485:2016,			